

CALIBRATION/INSPECTION CERTIFICATE

Specialty Biomedical certifies that this instrument, having been tested and calibrated, meets its published specifications. Specialty Biomedical certifies that the standards against which this instrument was calibrated are on record and that all measurements are traceable to National Institute of Standards and Technology.

Date	26-Oct-20	Technician	John P. Sample
<b style="font-size: 1.2em;">SPECIALTY BIOMEDICAL			
Data Control ID		SMPL-1000	
Device Type		VSM	
Manufacturer		GE	
Model		Procure 400	
Serial No		QW12345TY	
Location		Nurse Station	
Next Due Date		10/26/2021	
2151 Harvey Mitchell Pkwy S. Suite 207 College Station, TX 77840 979.764.2706			
Customer ID		SMPL	
Customer		Sample Customer	
Address		1234 Main St Anytown, TX 77845	

Blood Pressure mmHg			
Set	Low	High	Measured
190	187	193	190.1
120	117	123	121
80	77	83	81.2

PulseOx Percentage %O2			
Set	Low	High	Measured
85	82	88	85
95	92	98	95
98	95	100	98

PulseOx Heart Rate BPM			
Set	Low	High	Measured
40	38	42	40
80	78	82	80
140	138	142	140

Measured temperature °F			
Set	Low	High	Measured
98.9	98.4	99.4	99.3

Test Equipment Used		
Description	Serial Number	Calibration Due
VS Vital Signs Analyzer	3743	5/1/2021
PO Pulse Oximeter Simulator	OX1734	5/1/2021

Safety Features	
General unit Condition	PASS
AC Cord Condition	PASS
Switches and controls	PASS
Display Condition	PASS

Comments	
-----------------	--

Technician Signature:

Inspection Results
PASS